

JUN 09 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130290</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Troy Clark</u>	ID# <u>4934</u> Date <u>6-6-10</u>
A Agency <u>Anchorage Police Department</u> Phone # <u>786-8900</u>	
Instrument Location <u>Jail (North Room)</u>	
B Alco S/N <u>2E</u> Target Value <u>.080</u> High Pressure <u>1000</u>	
Alco Test Values <u>.079</u> <u>.079</u>	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> <u>CRB</u>	
6/11/10	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of June, 2010.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290 ✓

JUNE 06, 2010 ✓

OPERATOR'S NAME:

CLARK/TROY

OPERATOR'S NUMBER: 4934

SUBJECT'S LAST NAME:

V

SUBJECT'S FIRST NAME/MI:

V

O.L. #: V12345

DEPT/AGENCY: ANC1

CASE/REPORT: 10-12345V

TEST TYPE: V

ALCO TARGET VALUE: .080 ✓

ALCO S/N: 2E ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.14 in ✓
ALCO TARGET .080 23:58
BLANK TEST .080 23:59
INTERNAL STANDARD VERIFIED 23:59
ALCO TV 30.14 in .079 ✓ 23:59
BLANK TEST .080 00:00
SUBJECT SAMPLE .080 00:01
BLANK TEST .080 00:02
ALCO TV 30.14 in .079 ✓ 00:02
BLANK TEST .080 00:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290 ✓

JUNE 06, 2010 ✓

TIME 23:44

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.14 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()-./0123456789:;<=>?AB CDEF G
HI JKLMNOPQRSTUVWXYZ[-_`abc defghijklmno
pqrstuvwxyz{|}~"